



The Aliph RadioVibrometer

USER'S MANUAL

Revision C Version 2

1 TABLE OF CONTENTS

1	TABLE OF CONTENTS.....	2
2	INTRODUCTION	4
2.1	GENERAL DESCRIPTION	5
2.2	SENSOR UNIT	5
2.3	CHARGER	8
2.4	NECK INTERFACE	9
2.5	PARTS LIST	10
2.6	GETTING STARTED	11
3	SPECIFICATIONS.....	12
4	PERFORMANCE MEASUREMENTS	13
4.1	TEST SETUP.....	13
4.2	PERFORMANCE TESTS	13
4.2.1	<i>Highpass Filter Distortion Compensation.....</i>	<i>15</i>
4.2.2	<i>Measurement sheet.....</i>	<i>16</i>
4.3	CALCULATING SNR & MINIMUM DETECTABLE AMPLITUDE	17
5	ACQUIRING ARV DATA.....	19
5.1	NATIONAL INSTRUMENTS' LABVIEW VIS.....	19
5.1.1	<i>Installation of the software</i>	<i>19</i>
5.1.2	<i>Main and configuration screens</i>	<i>20</i>
5.1.3	<i>Real time data display.....</i>	<i>21</i>
5.1.4	<i>4 channel recording.....</i>	<i>23</i>
5.2	RECORDING DATA FROM THE GLOTTAL AREA.....	25
5.3	RECORDING DATA FROM OTHER LOCATIONS.....	26
6	FILTER DISTORTIONS DUE TO HP FILTER.....	27
7	SAFETY & COMPLIANCE	29
8	FAQ.....	30

9	TROUBLESHOOTING.....	33
9.1	ARV SIGNAL	33
9.2	LABVIEW PROGRAMS	33
9.3	TECHNICAL SUPPORT	34
10	CONTACT INFORMATION.....	35

2 INTRODUCTION

Thank you for choosing an Aliph sensor. We hope this sensor will serve you well for many years and become the basis for many discoveries.

The Aliph RadioVibrometer (ARV) was developed at Aliph for the purpose of detecting vibratory motion of human tissue. The present sensor is the fourth iteration of a design first constructed at Aliph in mid-1999. It was designed to be mass-producible and easily constructed using standard parts.

The sensor is an extremely sensitive phase-modulated quadrature motion detector that accurately determines the motion vs. time of one or more moving objects in its field of view. If several moving objects are present and detectable, the ARV signal will represent a combination of their motions. Thus the output of the ARV is a one-dimensional voltage that represents a combination of three-dimensional velocities. Therefore, care must be taken when interpreting the ARV signal as it is not always clear what object is responsible for the observed signal.

This manual is intended to inform a scientifically minded user how to use the ARV to record vibratory signals associated with human speech. It is not necessary for the user to have a background in electromagnetics (EM) or radiofrequency (RF) to use the sensor effectively. Although the sensor can be used to detect any type of motion anywhere on the body, we will restrict ourselves to motion caused by voiced speech in the sub-glottal and cheek/jaw areas. As we will see, the subglottal region can yield information on the excitation function of human speech and the cheek/jaw area may contain phonetic information about the produced speech. The user is encouraged to capture signals from other parts of the body, but we will restrict this manual to the two above as they return the most reliable and repeatable signals.

Any comments on the sensor or this manual are welcome and contact information is included in chapter 10. Additional information on the sensor and its interaction with the

human body can be found in a white paper entitled “Theory and Use of the ARV”, available at our [web site](#).

2.1 General Description

The ARV consists of three major subsections – the sensor, the antenna, and the battery and its charger. The battery should be the only contact between the sensor and charger, as the charger is not to be used with the sensor. The rechargeable battery simply slides into a receptacle on the side of the sensor body or charger. On the opposite side on the sensor is a belt clip that can be used to attach the sensor to a belt or similar object. The antenna plugs into the top of the sensor body with the supplied reverse polarity SMA cable. Due to FCC regulations, only the antennae supplied by Aliph can be used with the sensor. If you wish to use other types of antennas, you should contact Aliph. Aliph is not responsible for any liabilities relating to user safety or FCC compliance that may arise from connecting antennas or other components not supplied and/or certified by Aliph.

2.2 Sensor Unit

The sensor body is shown in Figures 2.1 and 2.2. The power switch is located on the bottom of the sensor body, next to an external power socket. **DO NOT PLUG ANYTHING INTO THE EXTERNAL POWER SOCKET! IT IS FOR MANUFACTURING PURPOSES ONLY AND SHOULD NOT BE UTILIZED BY THE END USER.** On top of the case are the antenna connector, the 3.5 mm audio stereo output line, a 20 dB gain potentiometer, and two indicator lights. The red power LED is always red and is lit when power is applied, while the battery LED varies:

Battery LED:	Steady Green	Steady Orange
Battery state:	Battery Good	Battery Low




Figure 2.1. Isometric view of the side of sensor body

Unlike previous revisions of the Aliph ARV sensor, there are no adjustable filters in Revision C. The filters are hardwired to a highpass 3-dB point of 20 Hz and a lowpass 3-dB point of 8 kHz. This should be sufficient to remove any unwanted noise in the signal. Further filtering may be accomplished after digitization of the signal.

As stated above, the Revision C sensor only uses a battery for power and cannot be powered from an AC source. It uses a 7.2 V, 1180 mAh camcorder battery, mounted on the side of the unit. Although an external power port is supplied, the end user should never use it. It is for production purposes and testing only. **CONNECTION OF ANY DEVICE TO THE EXTERNAL POWER ADAPTER CAN RESULT IN SERIOUS DAMAGE TO THE SENSOR AND IS NOT COVERED IN ANY WARRANTY, EXPRESSED OR IMPLIED.** It is recommended that users keep an extra battery on hand for experiments intended to last more than 5 or so hours. Then one battery may be charged while the other is being used.




Figure 2.2. Isometric view of the top of the sensor body.

The gain potentiometer should be used to control the signal size so that the size of the signal corresponds well to the A/D recording device. For example, if the A/D is set to acquire signals from ± 1 Volt, the gain should be set during normal voiced speech to approximately ± 0.5 V. This allows for larger transients to be captured while using most of the resolution of the A/D. Setting the gain to be too large can result in clipping of the signal, where the level of the signal no longer corresponds to the amplitude of motion. This distorts the signal and should always be avoided by using the potentiometer to reduce the gain.

On the top of the sensor near the potentiometer are the antenna connector and the stereo output connector. The stereo output contains the demodulated in-phase (I) and quadrature (90 degrees out of phase, or Q) signals. These signals represent the motion of the target and it is recommended that the largest energy signal of the two be used, as it is

normally the best (least distorted) representation of the motion. The smaller the signal, the more distortion can be present. In any situation, however, one of the modulations will always be nearly distortion-free due to the use of quadrature modulation.

The antenna connector is used to connect the Aliph antennae to the reverse-polarity cable supplied by Aliph. Due to FCC requirements, it is not possible to use antennae not supplied or approved by Aliph. Only the supplied antennae are to be used with the supplied reverse-polarity cables. If there is a specific antenna that you wish to use with the Aliph sensor, please contact us.

2.3 Charger

The charger will charge a battery in its battery bay ONLY. Simply slide a battery into the charger bay and plug in the AC adapter to begin the charging sequence. The charger has two LEDs – one red LED to indicate that the AC power is connected, and red/green LED to indicate charging state. The charging state is indicated as follows:

LED:	Steady Red	Red with slow flash	Red with rapid flash	Steady Green
Charging state:	Normal charging	Minor fault, slow charging	Major fault, follow reactivation procedure	Recharging complete

“Slow flash” means 1 flash every 5 to 10 seconds, “rapid flash” is 1 flash every second.

A fault may be introduced in the battery if it is deeply discharged.

To reactivate a battery with a major fault (deeply discharged or shorted) follow this procedure:

Place the battery in the charger and power up the charger. Allow the battery to remain in the charger for at least 10 minutes with the Charge Mode LED flashing quickly. After more than 10 minutes, remove the battery and disconnect the charger from AC power. Re-insert the battery into the charger and power the charger unit. If this procedure does


not reactivate the battery (i.e.. the charging state LED is still flashing quickly) the battery has been destroyed and a new battery must be procured.

2.4 Neck interface

The near-field neck interface (a more general term than antenna, and more appropriate in this case) supplied with the sensor is specially designed to work well next to human tissue. Normally, antennae are designed to function efficiently in air, which has a relative dielectric constant near unity and very low conductivity. However, at 2.4 GHz, the body has a relatively high dielectric constant (ranges from ~10-50) and significant conductivity (~2 S/m). Operation of the antenna near the body therefore affects the efficiency and resonant frequencies of the antenna and this must be taken into account during the design and testing phase.

The net effect of the operation of the antenna near the body is the destruction of the far-field structure of the antenna. Thus, a near-field device is needed. The neck interface supplied with the ARV is designed to concentrate the near field in a known area so that the area of interest can be targeted.

The near-field neck interface is shown in schematic representation below. It has proven to work reliably on a variety of subjects without tuning or customization. Antennae for use in other locations (such as the nose or cheek) are available from Aliph.



Schematic representation of the new Aliph ARV neck interface.

2.5 Parts List

The parts that were included with your Aliph ARV kit should include the following:

Quantity	Description
1	Revision C ARV Unit
1	Near-field neck interface
1	Reverse polarity double-shielded SMA cable
1	3.5 mm Stereo to RCA Cable
2	RCA to BNC connectors
1	Battery charger with AC power cable
1	Li-Ion rechargeable battery
1	CD with Data Acquisition software and support materials
1	User's manual

Please take a moment to check the contents of your kit and contact us if there are any discrepancies.

Additional batteries and chargers are available and can be ordered directly from Aliph. Only the antennae supplied by Aliph have been approved with by the FCC as legal to use with this device. If you would like to use an antenna of your own design with the ARV, contact Aliph and we will make arrangements for the necessary manufacturing and testing procedures.

2.6 Getting Started

- Step 1: Charge the battery if needed
- Step 2: Slide the battery provided into the battery bay on the sensor body
- Step 3: Firmly attach the double-shielded SMA cable to the operating antenna port
- Step 4: Firmly attach the near-field neck interface to the SMA cable
- Step 5: Insert the stereo 3.5 mm cable between the sensor and a DAQ device (such as an oscilloscope)
- Step 6: Switch on the sensor unit (switch located opposite the antenna connector)
- Step 7: Apply the antenna to your trachea (and speak) to observe signals
- Step 8: Adjust the audio output gain using the sensor gain adjustment knob

Note: Please read instructions on proper battery charging routines, antenna uses, and DAQ guidelines.

3 SPECIFICATIONS

The ARV sensor is a continuous wave phase modulated radiofrequency device that operates at a frequency of precisely 2.4 GHz. It uses a monostatic antenna configuration that uses reverse polarity SMA connectors. The maximum theoretical RF power output is 1 milliwatt while the effective power will depend on the efficiency of the antenna. It returns a voltage signal limited to +/- 2.5 V that indicates the amplitude of motion of a periodic or quasi-periodic motion of a target in its field of view. It is capable of measuring very small (micron order) vibrations from 20 Hz to 8 kHz both inside and outside the body. It operates on a rechargeable 7.4 V Li-Ion battery and draws approximately 170 mA of power.

ARV	
Maximum RF Power Output (Antenna Port)	0 dBm (1 mW)
Audio Signal Output	~5.0 V max peak-to-peak
Output Signal Operating Bandwidth	20 Hz - 8000 Hz
Maximum Sensitivity In Air (at 10dB SNR)	~5 μ m
AC Power Supply	110V - 120V AC
Input Voltage	6V - 9V DC (Regulated)
Operating Temperature (recommended)	0° to 35° C
Dimensions (approx)	100 x 70 x 20 mm
Weight	185 g

Battery	
Battery Capacity	1180 mAh
Battery Maximum Output Voltage	DC 8.4 V
Low Voltage Threshold	6 V
Operating Temperature	0° to 40° C
Dimensions (approx)	55 x 40 x 20 mm
Weight	76 g

4 PERFORMANCE MEASUREMENTS

The performance criteria for this type of device are defined by its operating characteristics and ability to directly represent the vibrational state of its targets. In this section we will explore the criteria used to determine the performance of the sensor.

The ARV is, simply put, a vibration analysis device. It can detect very small quasi-periodic or periodic motion of targets with very little distortion in one or both of the quadrature-related channels. As the ARV is a vibration detector, it is reasonable to use a NIST-traceable shaker table to check the accuracy of the ARV.

4.1 Test Setup

We chose for our testing the Gilchrist Technology Model 4000 portable shaker table. Information on this table can be found at <http://www.gilchristtechnology.com>. The target was a 4 inch diameter aluminum disk, manufactured to be under the 500 gram weight limit for the shaker table. A calibrated accelerometer was included as part of the table, and the frequency and amplitude of the shaker signal can be varied from 20 Hz to 10 kHz with a maximum displacement of 0.2 mm peak-to-peak at 100 Hz. The accelerometer output was available and was compared to the sensor outputs.

For the tests, a 2.45 GHz antenna (one that is efficient in air) was connected to the ARV of interest and placed in a test stand above the shaker head and lowered to just above the shaker table. Although the antenna is usually isotropic, normally the only vibrations in the correct frequency range are those of the shaker head, so interference was not a problem.

4.2 Performance Tests

The signals from the ARV and the accelerometer signal were recorded using the supplied Aliph ARV DAQ software and an analog to digital data acquisition system recording at 20 kHz. The distance measurements began at 0.8 cm ($\sim\lambda/16$) and were increased by $\lambda/16$

twice. At these distances the first measurement is taken with the I channel near its maximum, the second when the I and Q channels are about the same size, and the third when the Q channel is at its maximum. This allows us to test all of the signals at their strongest and weakest points. There was occasionally increased noise in the signals due to the proximity and perfect reflectivity of the metal target which was not observed when the antenna was placed on the skin. This caused the phase measurements to vary somewhat, but this should not be a problem in normal use.

The frequency of the shaker was set at 75 and 150 Hz, with a few seconds between recordings to allow the head to stabilize. These frequencies are slightly lower than normal male and female pitch, respectively, and were chosen mainly because above these frequencies the amplitude of the shaker head diminished rapidly to the point where accurate (± 5 degrees) phase measurements could not be made. Two different shaker amplitudes at each frequency were recorded and compared to the relative change in magnitude of the largest ARV signal to test the linearity of the ARV. After processing (see below), the accelerometer and ARV signals were compared to ensure that the phase and frequency information was correct.

To compare the position measured by the sensor to that derived from the accelerometer, we can either differentiate the sensor signal twice to get acceleration or integrate the accelerometer signal twice to get position. Or, since the accelerometer signal is composed of only a single frequency:

$$a(t) = A_0 \sin(2\pi ft + \phi)$$

the actual position signal will also be a single frequency

$$x(t) = X_0 \sin(2\pi ft + \phi)$$

When we differentiate this twice with respect to time, we get

$$x''(t) = a(t) = -(2\pi f)^2 X_0 \sin(2\pi f t + \phi).$$

Setting the amplitude factors equal, we see we may derive the position signal from the accelerometer signal quite simply:

$$x(t) = \frac{-A_0}{(2\pi f)^2} \sin(2\pi f t + \phi)$$

This requires only a sign change (180 degree phase shift) and division by $(2\pi f)^2$. We are not, however, ready to compare the signals directly – we still need to account for the phase distortion of the inherent ARV 20 Hz HP filter.

4.2.1 Highpass Filter Distortion Compensation

The ARV has a built-in HP filter with a 3-dB point of about 20 Hz, and several other selectable filters to allow the user to examine several frequencies of interest. These filters change not only the magnitude but the phase of the signal that passes through them, distorting the resulting filtered signal. Therefore if we want to compare the accelerometer and ARV signals directly, we can either attempt to remove the filter effects from the ARV data, or we can filter the accelerometer signal with models of the ARV filters that have similar magnitude and phase responses. The second method is much simpler and is what we will employ.

The measured response of the 20 Hz filter is shown in Chapter 6 in Figures 6.1 and 6.2. It is a relatively simple 2nd order analog highpass filter coupled with a delay of about 150 μ sec. The only real distortion is caused by the highpass filter below about 300 Hz. Above that, the phase is linear. To compensate for this phase distortion in our tests the accelerometer data was filtered with a model constructed using Matlab. This compensation assures that both the accelerometer and sensor signals have been distorted in the same manner, making a meaningful comparison possible. Once the differentiation and highpass filter distortion compensation operations were performed on the accelerometer signal, the result was compared directly with the sensor signals.

4.2.2 Measurement sheet

After the data at each frequency, antenna-to-target distance, and amplitude was recorded it was processed to determine the following:

1. Dominant ARV channel (dG)
2. Frequency of vibration as recorded by the accelerometer (Hz)
 - a. Calculated using a PSD, 0.1 Hz resolution
3. Frequency of vibration as recorded by the dG (Hz). This should be exactly the same as that recorded by the accelerometer.
 - a. Calculated using a PSD, 0.1 Hz resolution
4. Strength of the fundamental signal compared to the 1st harmonic (dB). This is an indication of distortion and 30 dB or greater is preferred.
5. Signal to noise ratio (SNR, dB). This should be 30 dB or more at every distance and frequency. The ARV normally is around 40-50 dB but can go as high as 90 dB.
6. Calculated shaker amplitude (mm, peak-peak). This is the amplitude calculated from the accelerometer signal given the accelerometer frequency of vibration from step 2. This did not always match up with the amplitude set on the shaker, as its accuracy can vary +-3%.
7. ARV RMS-derived amplitude (V, p-p). The RMS amplitude of the ARV was calculated in Volts and then transformed to peak-peak.
8. Minimum amplitude detectable (μm , p-p). See below for details. A measurement independent of target amplitude that describes the sensitivity of the ARV. 1-10 μm is good, 0.1-1.0 μm is great, and $< 0.1 \mu\text{m}$ is superlative.
9. Phase difference between the filter-compensated accelerometer signal and the dG signal (degrees). We would like to see differences of less than +-20 degrees, although there are some cases where we observed slightly more than that, usually at the lower frequencies. For the most part, though, even a difference of 20 degrees is negligible and should not affect the performance of the device on the skin. The reasons that there is some variation in the phase are:
 - a. Slight differences in the components that make up the 20 Hz HP filter and the 8000 Hz LP

- b. Noise due to the large reflected signal of the metal disk
- c. Poor model fit

10. Linear amplitude error (percent). This is calculated at the same distance and frequency but for different target amplitudes using the information from Steps 6 and 7. If the change in amplitude is perfectly linear, the error should be 0%. We are looking for errors less than 10%. Possible reasons for error are slight differences in distances, rounding errors, and low amplitude resolution. For example, we can only measure the shaker amplitude reliably to 3 significant figures (i.e. 0.095 and 0.191 mm). We can also only measure voltage to 2 significant figures (i.e. 0.80 and 1.58 V).

$$\text{error} = \frac{1.58 - \left(\frac{0.191}{0.095} \cdot 0.80 \right)}{\frac{0.191}{0.095} \cdot 0.80} \cdot 100 = 1.8\%$$

An error in the ARV measurements of only ± 0.02 V can increase the error to 5.4%, and ± 0.04 V can lead to an error of 8.9%. Therefore, considering the accuracy of our system, $\pm 10\%$ is reasonable.

4.3 Calculating SNR & Minimum Detectable Amplitude

At each test frequency the SNR was calculated using a background noise file (no shaker motion) as the noise level. The result was then extrapolated back to what the signal level would be at 10 dB SNR (amplitude of the signal would be 3.16 times the amplitude of the noise). This allows us to determine the minimum amplitude that can be reliably detected at that frequency. Of course, this test is in air, so any extrapolation to vibrational amplitudes in tissue will have to account for conductive losses, imperfect reflections, and scattering losses. The minimum amplitude for which a 10 dB SNR is possible is calculated by

$$\text{min_amp} = S_0(\text{mm}) \frac{3.16(N_1(\text{V}))}{S_1(\text{V})}$$

where S_0 is the observed accelerometer amplitude in mm, N_1 is the noise amplitude in volts, and S_1 is the signal amplitude in Volts. We may use either peak-to-peak or RMS voltages in the amplitude calculations, as we are dealing with ratios of amplitudes. For

example, if the observed voltage for an amplitude of 0.1 mm was 2 Volts, and the background noise amplitude was 20 mV, then the initial SNR would be

$$\text{SNR}_0 = 20 \log\left(\frac{2000}{20}\right) = 40 \text{ dB}$$

and the smallest detectable signal with an SNR of 10 dB would be

$$\text{min_amp} = 0.1 \text{ mm} \frac{3.16(0.02)}{2} = 0.003 \text{ mm} = 3 \text{ } \mu\text{m}$$

which corresponds to a voltage of 60 mV, three times that of the noise level of 20 mV. This measurement is a convenient way to view the SNR independent of target amplitude.

5 ACQUIRING ARV DATA

The ARV is easy to use – just place the antenna and examine the output. However, it requires some practice and finesse to operate correctly and robustly. The sensor itself is usually not at fault, it is normally the combination of antenna placement and cable movement that give new users the most problems. The cables will contribute noise to the system if they are moved or vibrated during data acquisition. If your subject is to move during the experiment, make sure that the cables are secured and not allowed to move freely. This will minimize the noise in the ARV signal.

5.1 National Instruments' Labview VIs

To assist users in visualizing and recording ARV data, stand-alone and regular Labview virtual instrument files are included as part of the ARV package. These files can display the ARV data plus two more data channels and record them in a binary or text file. To use the software, data acquisition hardware and software from National Instruments (the E series at <http://sine.ni.com/apps/we/nioe.vp?cid=1038&lang=US> is recommended, it uses NI-DAQ software) must be installed and correctly configured. To test your configuration after installation, try the “Test Panel” option using the NI Measurement and Automation Explorer.

This software is provided as a courtesy by Aliph and has only been tested on National Instruments hardware running under Windows 2000 and Windows 98.

5.1.1 Installation of the software

The installation routine for the executable software is on the CD. Simply place the CD in your CD drive. If the setup screen does not automatically start, open the CD and select “setup”. The executable files will be copied into a directory of your choice, and a directory called “Aliph ARV DAQ” will be created on your Start→Programs menu. Clicking on “Aliph ARV DAQ” will launch the program. Remember that you must have the NI hardware and software installed before you can use the Aliph software.

As a courtesy to those that may wish to customize the VIs, the source code is included in the \source directory. To use these files, you must have at least the Professional version of Labview 6.1 installed on your computer. Otherwise you may use the stand-alone applications described above.

5.1.2 Main and configuration screens

In Figure 5.1, the main menu of the Labview program is shown. Before data can be recorded or displayed, the hardware must be configured. This is why some of the buttons are disabled (grayed out) – because in this screen the hardware has not yet been configured. To configure the hardware for the first time, press “Configure DAQ” to open the hardware configuration screen. If you have saved a configuration default or other



Figure 5.1. Main screen of the Aliph data acquisition package.

file, you can press “Load config default” or “Load config file” to save time. You must load a config file or manually enter a configuration every time the program executes.

The configuration screen is shown in Figure 5.2. Here, you can load the default setting (Load default), load any configuration file (Load...), save the current settings as the default setting (Save as default), save the current settings to a file (Save...), do nothing and leave the program (Cancel), or use the current settings for the current DAQ session without saving them to a file (OK). The configuration files can be named anything; they are stored in text format. A help menu explains all the settings available.




Figure 5.2. The hardware configuration screen.

Hitting “OK” or “Cancel” will return the user to the main menu. If a valid configuration has been chosen, the previously unavailable options will be enabled. In the future, you may load the default file or any saved configuration file from the main screen and skip the manual configuration entry.

5.1.3 Real time data display

Two options are given for real-time data display – a 1 channel display and a 4 channel display. They work in the same way, so we’ll go through just the 1 channel, shown in




Figure 5.3. Real time configuration screen.

Figure 5.3. First, select a channel to display and a time scale to display it with. Values from 5 msec to 1000 msec may be chosen, and that amount of data will be displayed at once. This value also controls the refresh rate. The scan rate in Hz is displayed below the time scale, and is set in the hardware configuration screen (see Figure 5.2). The scan rate is the entered scan rate divided by the decimation factor.

Press “Start” and the screen will display the channel selected. If the time scale is grayed out, it cannot be changed when data is being acquired. When the display is no longer needed, press “Stop” and then the data that was being displayed is available to view. Use the palette on the right side of the plot to zoom in and out of the data. When finished, press “Start” again to restart the data display, or press “Return” and the window will close and you will be back at the main program (Figure 5.1).

The 4 channel real time display system works the same way, except that all four channels (from 0-3) are displayed simultaneously. The sampling rate is again set in the configuration screen.

5.1.4 4 channel recording

The final algorithm offers 4 channels of data recording in either binary or text format. All four data streams are stored in the same file. The binary is recommended due to its smaller size, and a Matlab file is included (see the readme.txt file on the CD) that will translate the .bin file to a Matlab matrix. The interface is shown in Figure 5.4. To begin, enter the filename of the file you want to save the data in and the amount of time you want to record (a setting of zero here will record until you press “Stop”). The type of file recorded will be determined by the “Data file format” drop down menu. If you have specified a file suffix (like .bin) in your filename, this will override the selection on the drop down menu. If you specify a three letter extension in your filename but it is not .bin or .txt, the program will return an error and terminate. The simplest thing to do is to just specify a filename without suffix and choose the type of file you want to save from the drop-down menu.

Now press “Start”, and the “Recording” light will come on and the program will begin to record the data. You can stop the data acquisition at any time by pressing “Stop”. When the data acquisition is complete, the “Recording” light will go out and the data will be displayed. Displaying the data can take a while if the file is long. You will now be able to examine the data in greater detail using the graph palettes on the right side of the graphs. When you are satisfied with the data, you may take more data by pressing “Start” again (be sure to change the filename first if you want to save the data!) or you can return to the main menu by hitting “Return”.




Figure 5.4. 4 Channel Data recording screen.

5.2 Recording Data from the Trachea

The glottis is the area between the vocal folds, and has been the subject of most ARV research until now (see Chapter 2). The largest ARV signal can be obtained by placing the neck interface directly under the chin on the laryngeal notch (Adam’s apple). The resulting signal represents a combination of tracheal wall vibration and anterior vocal fold motion. For more information, see the white paper “Theory and Use of the ARV” available at [our web site](#).

A tracheal vibration signal without the vocal fold contribution can be detected lower down on the trachea. However, the signal is smaller and contains fewer harmonics due to tissue damping and the receding of the trachea into the neck as it approaches the chest.




Figure 5.6. Example of an ARV signal sampled at 10 kHz.

5.3 Recording Data from Other Locations

The ARV is sensitive enough to pick up vibrations from a number of places on the body, including the back and sides of the neck and the chin and jaw. Vibrations from different areas can contain different information – for example, the glottal area signal reflects the pressure pulses that become the speech, and the cheek area contains more information about the sound being produced. Experiment with the sensor and the interface, and you may be surprised at what you find! Aliph constructs customized interfaces for different applications. Contact us for more information regarding a custom interface.

6

FILTER DISTORTIONS DUE TO HP FILTER

The ARV electronics includes a simple analog HP filter to reduce low frequency noise. This filter has a 3-dB point at approximately 20 Hz. The measured response of the filter is shown in Figures 6.1 and 6.2. It is clear that there is some non-linear phase distortion of the filtered signal for frequencies below about 300 Hz. Above that frequency, the only phase response is a linear phase, which translates directly to a simple delay. In this case, the delay is about 150 μ sec. This delay does not in any way affect the shape of the ARV signal, and therefore does not need to be removed. For those desiring absolute accuracy for signals with components greater than 300 Hz, simply advance the ARV signal by 150 μ sec. However, remember when comparing the ARV and acoustic signals that the ARV signals are traveling at the speed of light and the acoustic signals at the speed of sound. A pressure pulse produced at the glottis will have to travel about 20 cm to get to a close-talk microphone near the lips, and at 330 m/s that will take about 600 μ sec. For all intents and purposes, the ARV signals are generated and reflected instantly. Adding in the 150 μ sec delay means that for proper alignment of the acoustic and ARV information, the ARV should be delayed approximately 450 μ sec. In practice this is rarely an issue.

Removing the nonlinear distortion of the 20 Hz filter for the frequencies of interest (~40 Hz and up) can be done, but will not be covered in this manual. An example is given in Burnett's thesis (see Chapter 2), but will not be needed for the vast majority of applications, as the phase distortion is insignificant for most frequencies and does not in any way affect the frequency content of the signal. The 100 Hz HPF option, however, can cause significant phase distortion and should not be used if it is not needed. Once again, it is recommended to select the 20 Hz and 8 kHz filter options for minimum phase distortion.

If you feel that you must remove the phase distortion from the 20 Hz filter, contact us and we will supply you with a Matlab program that will be able to remove a significant

amount of the distortion. Perfect removal is not likely due to normal variation in the filter response.




Figure 6.1. Magnitude response of the intrinsic 20 Hz filter of the ARV. The frequency range here is from DC to 500 Hz. The spike near DC is not real, but rather an inaccuracy of the instrument.




Figure 6.2. Phase response of the intrinsic 20 Hz ARV filter. The rapid change at 3337 Hz is simply the phase wrapping around to stay between ± 180 degrees. This phase response has the characteristics of both an IIR and FIR filter. The frequency range is from DC to 10000 Hz.

7 SAFETY & COMPLIANCE

Your ARV device is a transmitter and a receiver. In operation, it transmits less than 1 mW (0 dBm) of power in the frequency range of 2400-2500 MHz. Your ARV device is designed to operate in compliance with the RF exposure guidelines and limits set by the national authorities (FCC) and international health agencies when used as instructed in this manual.

The ARV complies with Part 15 of FCC regulations concerning the use of intentional and unintentional radiators in the 2.4GHz ISM band.

For additional information concerning interference or safety issues, please refer to the “Aliph ARV Safety and Compliance Report”, which will be provided to ARV owners upon request.

Does the ARV “image” the vocal folds?

No. The ARV is a vibration detector and simply returns a signal that is proportional to the movement of a target in the area. If there are several vibrating surfaces, the resulting signal will be a nonlinear combination of all of the movements of the surfaces. Since the ARV only detects signals above about 20 Hz or so and the skin depth (depth in tissue that the EM wave is reduced in energy to $1/e$) is only a few cm, normally there will only be one major contributor to the signal. However, care must be taken when interpreting the signal.

Is the ARV safe to use?

The ARV is a radiofrequency (RF) device that emits electromagnetic (EM) energy at a frequency of 2.4 GHz, the same frequency many cordless phones and wireless headsets use. However, the ARV emits less than 1/1000 of the power of a mobile phone and complies completely with all US and foreign standards for continuous exposure to EM energy at this frequency. Future versions of the ARV will emit even less energy.

Does the antenna have to be touching the body to get a signal?

No. The use of normal, in-air antenna held off the skin can return a small but useable signal. However, for this study the emphasis has been for on-body antenna to reduce the power output as much as possible and to make the signal as robust and repeatable as possible.

Does the ARV directly sense the motion of the vocal folds?

Yes, but only a small part of the anterior portion. The movement of the folds disturbs the near-field of the interface, resulting in a signal from the ARV. This signal is combined with the signal normally detected due to the vibration of the trachea.

We don't have a data acquisition (DAQ) PC card, we only have a DAT tape player (or similar). Can we still use the Aliph Labview programs to collect data?

The Aliph DAQ program only supports PC-based data acquisition. We recommend any of the E-Series National Instruments DAQ boards (such as the PCI-6023E or the PCI-6034E) as well as their line of BNC based connector boards such as the BNC-2110.

They are not expensive and offer reliable performance.

What sampling frequency should we use to capture the ARV data?

Although the ARV allows signals as high as 8 kHz to be detected, the response of the tissue to the speech pulses is poor above 2-3 kHz and it is unlikely that any significant energy will be in the ARV above 4-5 kHz. Still, to ensure that aliasing is not present, we have found that recording the data at 24 or 32 kHz and then digitally filtering and decimating to 8 kHz has been quite successful in capturing the ARV signal without aliasing. In a pinch, 16 kHz will do, but is not recommended.

Does the ARV sensor work on all people?

We have been able to get a useful subglottal signal from all of our subjects. This does not mean that all people will return one, but the vast majority of people should yield a useable tracheal signal. However, the cheek signal is not as reliable. If the subject is a soft speaker or if the tissue of the cheek is incompatible with our antenna, a useful cheek signal is sometimes not available. We are working on a more sensitive version of the ARV that will ensure a cheek signal for all subjects.

Does the tracheal ARV signal contain phonetic information?

No; or at the most, very little. The subglottal pulses detected by the ARV do not contain significant phonetic information. Since the phoneme is generated by the rearrangement of vocal articulators in the supraglottal region (including the larynx, pharynx, and oral cavity) and the reflection of acoustic energy at every change in cross-sectional area in the vocal tract, there is undoubtedly some phonetically shaped air pressure leakage from the supra- to the subglottal region. However, it will be quite small, as the glottis (the air space between the vocal folds) is small compared to the vocal folds and most energy will

be reflected back to the vocal tract. In addition, the large pulses of air that excite the vocal tract will be many times stronger than the small amount of leakage past the folds. That is not to say that with concentrated effort some phonetic information may not be obtained.

What about the cheek/jaw information? Is that phonetically related?

Yes. The vibrations of the cheek and jaw are caused by oral cavity air pressure fluctuations, which are the result of the shaping of the glottal pulses by the vocal tract. Depending on the phoneme, the vibrations should have different frequency content and amplitude. The frequency content comes from the frequency content of the phoneme being produced, and the amplitude variations are related to how well the phoneme being produced couples with the air. An “ee”, for example, does not couple well with air and there is ample energy inside the oral cavity to vibrate the cheek and jaw. An “ah”, on the other hand, couples very well with air and there is little oral cavity energy left to vibrate the cheek and jaw.

9 TROUBLESHOOTING

9.1 ARV Signal

The ARV is a relatively simple device to use, especially in the subglottal area. Try that area first, as that is the area where the signal is the largest and most robust. If there is still no signal from the subglottal area, try the following troubleshooting guide:

- Check the gain potentiometer to make sure it is all the way up (full clockwise).
- Ensure that the battery is charged (check the “battery indication” light).
- Check the antenna connection to make sure it is firmly connected to the sensor.
- If using a rectangular patch antenna, make sure the long edge of the patch is perpendicular to the trachea.
- Try a different antenna and make sure that it is flat against the skin.

If a signal is returned from the trachea, but not the cheek, try the following:

- Use the spiral antenna. It has the potential to match better with the cheek.
- Vary the position of the spiral antenna while you make the sound “mm” with your mouth closed. This should result in the largest signal amplitude possible.
- Vary the amount of pressure used to press the antenna to the cheek. Vary it from lightly touching to firmly touching.
- Try to match the skin better using cellophane tape, as described in Chapter 5.

9.2 Labview Programs

Since we are not supplying any data acquisition (DAQ) hardware with the ARV, support for DAQ hardware issues is not included in your package, but may be available on a pay-per-time basis. Support for problems you may be having with the Labview software is included, **BUT ONLY FOR SOFTWARE THAT HAS NOT BEEN MODIFIED**. If you

modify the included Labview VIs for your own use, as is allowed, you are responsible for the results. Don't come crying to us, we have enough problems of our own.

Before calling us with software questions, you must make sure your hardware system is working correctly by using the Test Panel in the Measurement and Automation Explorer, part of the NI-DAQ software series (available with any hardware purchase or download from www.ni.com). Input a signal from a synthesizer or frequency generator on the channels that you want to use and check to make sure that they are being correctly received by the DAQ card. Common mistakes include connecting the incoming signals to the wrong DAQ inputs and not ensuring tight cable connections. The Aliph software uses channels 0, 1, 2, and 3 to input data. It is not necessary to use all of these channels, but the signals you want to record should be connected to one of the above four channels.

If you are detecting these known signals on the correct channels, then the next step is to see if the ARV is producing a signal. You can use an oscilloscope or input to a computer soundcard to see if ARV signals are being produced from the subglottal region. The subglottal signal can be recorded and sounds like muffled speech with no phonetic information. If these signals are being recorded correctly, and the input to Labview is working correctly, the system should function.

9.3 Technical Support

If you are still having problems, please feel free to contact us. Five hours of support time, which should be scheduled in advance, are included with the purchase of a new sensor. Contact us and we'll be glad to go through your problem with you. After the included support time is exhausted, we would be happy to provide more support although this may result in additional fees.

10 Contact information

You may contact Aliph in the following ways:


Phone: 650-276-0515

Fax: 650-244-6465

Email: gburnett@aliph.com

Mail: 1485 Bayshore Blvd.
Suite 382, Mail Stop 93
San Francisco, CA 94124

Aliph ARV Quick Reference



Quick start:

- Step 1: Charge the battery if needed
- Step 2: Slide the battery provided into the battery bay on the sensor body
- Step 3: Firmly attach one SMA cable to the operating antenna port
- Step 4: Firmly attach 1 antenna (e.g. 9p6+7_80) to the SMA cable
- Step 5: Insert the stereo 3.5 mm cable between the sensor and a DAQ device (such as an oscilloscope)
- Step 6: Switch on the sensor unit (switch located opposite the antenna connector)
- Step 7: Apply the antenna to your trachea (and speak) to observe signals
- Step 8: Adjust the audio output gain using the sensor gain adjustment knob

Note: Please read instructions on proper battery charging routines, antenna uses, and DAQ guidelines.

DO NOT USE THE EXTERNAL POWER PORT! THE ARV CAN BE SERIOUSLY DAMAGED!